

S-O.No. 21000820819253 d/h, 4/3/20



**Contract**

Contract No: GEMC-511687749205259

Generated Date:04-Mar-2020

Bid/RA/PR No:GEM/2019/B/47 1654

Organisation Details		Buyer Details	
Type:	Central PSU	Name:	Ashish Sewakram Nagarkar
Ministry:	Ministry of Coal	Designation:	Dy. Manager
Department:	Materials Management	Contact No.:	0712-2511381-5529
Organisation Name:	Western Coalfields Ltd	Email ID:	asnagarkar@westerncoal.gov.in
Office Zone:	Western Coalfields Limited / Nagpur	GSTIN:	27AAACW1578L1ZW
		Address:	1st Floor, MM Wing, WCL Hqrs, Coal Estate, Civil Lines, NAGPUR, MAHARASHTRA-440001, India

Financial Approval Detail
IFD Concurrence : No
Designation of Administrative Approval : Director (Technical)Projects and Planning
Designation of Financial Approval : GM(Fin)/c-1

Seller Details	
Company Name:	MODITECH INTERNATIONAL
Contact No.:	9216474081
Email ID:	moditech1@gmail.com
Address:	4A METRO PLAZA SCO 54 55,SECTOR 9 D,SECTOR-9D, CHANDIGARH, CHANDIGARH-160009, -
MSME verified:	Yes
MSE Social Category:	General
MSE Gender:	Male

Product Details							
#	Item Description	Category Name	Model	HSN Code	Ordered Quantity	Unit	Price (Inclusive of all Duties and Taxes in INR)
1	MULTIPARA MONITOR - LOW END	Multipara Monitor - Low End	MEDOXY MPM12	HSN not specified by seller	24	pieces	912,000
Total Order Value (in INR)							912,000

Consignee Detail				
S.No	Consignee	Item	Quantity	Expected Delivery Date
1	Asaari Mahender - Landline-- asaari.mahender@nic.in Regional Stores, WCL P.O. GHUGUS COLLIERY Distt Chandrapur, M.S ; Pin 442 505, CHANDRAPUR, MAHARASHTRA-442505, India	MULTIPARA MONITOR - LOW END	8	02-Jun-2020
2	Anand Kumar Pandey - Landline-6260565123- akpandey989.wcl@nic.in Regional Stores, WCL Pench Area P.O CHANDAMETA Distt. Chhindwara, M.P 480447, CHHINDWARA, MADHYA PRADESH-480447, India	MULTIPARA MONITOR - LOW END	8	02-Jun-2020
3	Rahul Rajendra Amane - Landline-07116-247181- rahul.amane@nic.in Regional Stores, WCL PO SILEWARA COLLIERY Dist. Nagpur M.S.; Pin 441 109, NAGPUR, MAHARASHTRA-441109, India	MULTIPARA MONITOR - LOW END	8	02-Jun-2020

Product Specification for MULTIPARA MONITOR - LOW END		
Specification	Sub-Spec	Value
Standards	Conformity to Certifications	CE European
	Conformity to Manufacturing Standards	ISO 9001 & ISO 13485
	Conformity to Safety Standards	ISO 80601-2-49 and IEC 60601-1 or BIS equivalent
Performance Parameters	Purpose	Low-end monitors are designed to perform basic vital-signs monitoring. These monitors may also be appropriate for IPD as well as OPD patients
	Type of Monitor	Modular
	Monitoring Parameters	ECG,HR, NIBP, PR, respiratory rate, SpO2, and body temperature
ECG parameters	Number of leads of ECG	5
	12-lead interpretation for ECG	Yes
	Arrhythmia detection and analysis	Yes
	Number of leads analyzed for ECG in single time	5
	Heart rate measurement range	At-least 30 -250 BPM
	Accuracy	± 5 BPM
	Minimum gradation	1 bpm
Respiration Parameters	Respiration Method	Impedance pneumography
	Display waveform of respiration	Yes
	Threshold control for respiration	Yes
	Respiration rate measurement range	At-least 0-100 bpm
	Minimum gradation of respiratory measurement	1 BPM
Pulse Oximetry Parameters	Measurement technique	Nellcor
	Other Measuring Techniques (In case measurement technique is nellcor, masimo , draeger, oxisure, trusignal put NA otherwise indicate the measurement technique)	NELLCOR
	Probe type supplied	Adult,Pediatric,Neonatal
	Type of adult probe	Reusable
	Type of pediatric probe	Reusable
	Type of neonatal probe	Reusable
	SPO2 measurement range	Atleast 40-70% & 70 to 99%
	Accuracy of SPO2 measurement	Better than ±1% for 40-70 range & better than ± 3% for 70-99 range
Minimum Gradation	1%	
NIBP Parameters	Measurement technique for NIBP	Step-deflation oscillometric
	Cuff supplied	Adult,Pediatric,Neonatal
	Cuff size for Adult	Medium
	Cuff size for pediatric	Standard
	Cuff size for neonate	Standard
	Blood pressure monitoring range	30-300mm/Hg
	Gradation	1 mm/Hg or better
	Number of channels for temperature	2