

S.O. 21100820819252 dtd. 4/3/2020



Contract

Contract No: GEMC-511687741072088

Generated Date:04-Mar-2020

Bid/RA/PR No:GEM/2019/B/483127

Organisation Details	Buyer Details
Type: Central PSU Ministry: Ministry of Coal Department: Materials Management Organisation Name: Western Coalfields Ltd Office Zone: Western Coalfields Limited / Nagpur	Name: Alok Kumar Designation: Asst Manager (MM) Contact No.: 0712-2512354-5585 Email ID: alokkumar@westerncoal.gov.in GSTIN: 27AAACW1578L1ZW Address: 1st Floor, MM Wing, WCL Hqrs, Coal Estate, Civil Lines, NAGPUR, MAHARASHTRA-440001, India

Financial Approval Detail
D Concurrence : Yes Designation of Administrative Approval : Director Designation of Financial Approval : GM Finance

Seller Details
Company Name: MEDICARE TECHNOLOGIES Contact No.: 07351210003 Email ID: amit.medicare@gmail.com Address: A 94 MMIG, SIKANDRA, Agra, UTTAR PRADESH-282007, - SME verified: Yes SE Social Category: General SE Gender: Male

Product Details						
Item Description	Category Name	Model	HSN Code	Ordered Quantity	Unit	Price (Inclusive of all Duties and Taxes in INR)
FOWLER BED WITH RAILLING	Patient Bed (Fowler and Semi Fowler)	KME-FB20	9402	12	pieces	197,988
Total Order Value (in INR)						197,988

Consignee Detail				
Sl No	Consignee	Item	Quantity	Expected Delivery Date
	Anand Kumar Pandey - Landline-6260565123- akpandey989.wcl@nic.in Regional Stores, WCL Pench Area P.O CHANDAMETA Distt. Chhindwara, M.P 480447, CHHINDWARA, MADHYA PRADESH-480447, India	FOWLER BED WITH RAILLING	4	18-Apr-2020
	Rahul Rajendra Amare - Landline-07116-247181- rahul.amare@nic.in Regional Stores, WCL PO SILEWARA COLLIERY Dist. Nagpur M.S.; Pin 441 109, NAGPUR, MAHARASHTRA-441109, India	FOWLER BED WITH RAILLING	4	18-Apr-2020
	Avinash Kumar - Landline-07172-237026- avinashkumar@westerncoal.gov.in	FOWLER BED WITH	4	18-Apr-2020

Product Specification for FOWLER BED WITH RAILLING

Specification	Sub-Spec	Value
Performance Parameters	Type of bed	Fowler
	Number of bed platform sections	4 for fowler bed
	Availability of Backrest section	Yes
	Backrest section linkage	welded with MS CRCA tube (14 G)
	Availability of Leg section	Yes
	Leg section linkage	welded with MS CRCA tube (16 G)
	Mattress platform shall be perforated	Yes
	Screw mechanism	welded with ERW MS tube (18 G) in M.S. cover made from (16 G) ERW tube
	Manual adjustments	Backrest , knee rest through two screw systems with thrust bearings individually manouvred by two handles for fowler bed
	Back rest - min 45 % of the frame length	Yes
	The bottom end of the H legs are provided with PVC shoes	NA in case of bed with castors
	The bed has head & foot panels detachable by hand without need of any tool	Yes
	Saline rod provision	four locations on the bed platform to hold stainless steel Telescopic Saline rod 12mm dia with 15.87 mm dia x 1.2 mm (18 G) stainless steel
	Provision for mounting of syringe pump	outer covering tube with a knob to mount syringe pump
	Finishing & workmanship shall be of high standard	Yes
	All corners shall be rounded off so that there shall be no sharp corners and holes should be burr free	Yes
	Coating for metal parts	M.S. tubular parts, linkages, flats are to be In-house, pretreated / shot blasted and Epoxy powder coated with coating thickness 50 to 100 microns
	Material of handle	SS
	Number of detachable moulded PP with PVC insert head and foot side panels	4
	Provision of castors	H' Type Legs with 600 mm Height with 125 mm(5") dia High Grade Synthetic Body Castors, two with brake, two w/o brake
	Number of hooks in Telescopic I V Pole	4
	Urine Bag Holder	Yes
	Moulded Chart Holder	Yes
	Provision of mattress	Yes
	Number of sections in mattress	4 for fowler bed
	Provision of collapsible Railing	Yes
	Overall Length of bed in mm with 5% tolerance	2050 to 2100
	Overall width of bed in mm with 5% tolerance	900 to 950
Size of Buffer in mm	100	
Bed Platform frame size with 5% tolerance	2070 mm L x 910 mm W	
Mattress Platform size with 5% tolerance	1970 mm L x 790 mm W	
Buffer to Handle Length with 5% tolerance	2275 mm	